

Certificate of Analysis

Product identification			
Product name	R-DiaRes		
Product reference	DDGR-46-L100	Manufacturing date	2020-10
Lot number	10000J	Expiry date	2022-10

Analysis procedure	
Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-R4-A100
Real Time PCR equipment (Supplier)	LightCycler 480II (Rohe)

QC Results				
Testing	Channel	Specification	Results	Status
Positive Control (Ct/Cp)	FAM	30.0 ± 3.3	29.1	Passed / Fail
Positive Control (Ct/Cp)	YD	30.0 ± 3.3	31.1	Passed / Fail
Negative Control (Ct/Cp)	N/A	Negative	Negative	Passed / Fail

Approval			
	Name & Function	Signature	Date
QC	COANET Anne QC Manager		23/12/2020
QA	Caroline TOUSSAINT QA Officer		23 DEC 2020

This batch has been released by Quality Assurance in compliance with Diagenode's Quality Management System.