

**Identification of the product**

Product name	R-DiaRes		
Product reference	DDGR-46-L100	Production date	2018-12
Lot number	2018L18/I	Expiry date	2020-12


**Analysis procedure**

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-R4-A100
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

**QC results**

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp) :	FAM	30.0 ± 3.3	28.6
Positive Control (Ct/Cp) :	YD	30.0 ± 3.3	30.4
Negative Control (Ct/Cp)	N/A	Negative	Negative

**Approval**

Name and function : <div style="border: 1px solid black; padding: 2px; display: inline-block;">                 Dominique d'Amico                  QA Officer             </div>	Signature: 	Date: 20-01-2019
---	--	---------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.