

**Identification of the product**

Product name	R-DiaLeg		
Product reference	DDGR-50-L100	Production date	2017-01
Lot number	03007A	Expiry date	2019-01


**Analysis procedure**

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-D2-R100
Real Time PCR equipment (Supplier)	LightCycler480 II (Roche)

**QC results**

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp) :	FAM	30.0 ± 3.3	28.2
Positive Control (Ct/Cp) :	YD	30.0 ± 3.3	28.0
Negative Control (Ct/Cp)	N/A	Negative	Negative

**Approval**

Name and function : D. Kusinida QA OFFICER	Signature: 	Date: 02/05/2017
--	--	---------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.