

Certificate of Analysis

| Product identification | | | |
|------------------------|--------------|--------------------|---------|
| Product name | R-DiaLeg | | |
| Product reference | DDGR-50-L100 | Manufacturing date | 2021-04 |
| Lot number | 07001D | Expiry date | 2023-04 |

| Analysis procedure | |
|------------------------------------|----------------------------|
| Volume PCR samples (µL) | 5.0 |
| Volume PCR mix (µL) | 20.0 |
| Master Mix reference | DMML-D2-R100 |
| Real Time PCR equipment (Supplier) | LightCycler 480 II (Roche) |

| QC Results | | | | |
|--------------------------|---------|---------------|----------|-------------|
| Testing | Channel | Specification | Results | Status |
| Positive Control (Ct/Cp) | FAM | 30.0 ± 3.3 | 29.8 | Passed/Fail |
| Positive Control (Ct/Cp) | YD | 30.0 ± 3.3 | 30.8 | Passed/Fail |
| Negative Control (Ct/Cp) | N/A | Negative | Negative | Passed/Fail |

| Approval | | | |
|----------|---------------------------------|--|-------------|
| | Name & Function | Signature | Date |
| QC | Philippe Nawol pc Specialist |  | 28-06-2021 |
| QA | Samira Lamani QA officer |  | 08 JUL 2021 |

This batch has been released by Quality Assurance in compliance with Diagenode's Quality Management System.