

Identification of the product

Product name	R-DiaLeg		
Product reference	DDGR-50-L100	Production date	2017-02
Lot number	2017B27//II	Expiry date	2018-02


Analysis procedure

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-D2-R100
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

QC results

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp) : <i>Legionella species</i>	FAM	30.0 ± 3.3	27.4
Positive Control (Ct/Cp) : <i>Legionella pneumophila</i>	YD	30.0 ± 3.3	28.3
Negative Control (Ct/Cp)	N/A	Negative	Negative

Approval

Name and function : D. MUSINDA QA OFFICER	Signature: 	Date: 07/03/2017
---	--	---------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.