

Identification of the product

Product name	R-DiaLeg		
Product reference	DDGR-50-L100	Production date	2017-10
Lot number	2017K06/I	Expiry date	2018-10


Analysis procedure

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-D2-R100
Real Time PCR equipment (Supplier)	LightCycler480 II (Roche)

QC results

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp) :	FAM	30.0 ± 3.3	29.9
Positive Control (Ct/Cp) :	YD	30.0 ± 3.3	30.0
Negative Control (Ct/Cp)	N/A	Negative	Negative

Approval

Name and function :	Signature:	Date:
<div style="border: 1px solid black; padding: 2px; display: inline-block;"> David Kusinda QA Officer </div>		22/11/2017

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.