

Identification of the product

Product name	R-DiaRhino		
Product reference	DDGR-60-L100	Production date	2019-07
Lot number	07009G	Expiry date	2021-07

Analysis procedure

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-R4-A100
Real Time PCR equipment (Supplier)	LightCycler480 II (Roche)

QC results

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp) : Rhinovirus	FAM	30.0 ± 3.3	28.4
Negative Control (Ct/Cp)	N/A	Negative	Negative

Approval

Name and function : S. Lombard QA Officer	Signature: 	Date: 1 oct 2019
---	---	---------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.