

**Identification of the product**

Product Name	R-DiaRhino		
Catalog Number	DDGR-60-L100	Production Date	2016-02
Lot Number	16146B	Expiry Date	2018-02

**Analysis procedure**

Volume PCR samples (µL)	5
Volume PCR mix (µL)	25
Master Mix reference (Supplier)	OPTIMA AB Master Mix 4x RNA (Diagenode)
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

**QC Results by Real Time PCR**

Testing	Specifications	Results
Positive Control (Ct/Cp) HRV	30.0 ± 3.3	FAM: 28.7
Negative Control (Ct/Cp)	Negative	Negative

**Approval**

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.

Approved by: A. ZALETELS  Date: 06 Apr 2016  
*QA/PC Manager*