

Identification of the product

Product Name	R-DiaRhino		
Catalogue Number	DDGR-60-L100	Production Date	2016-08
Lot Number	16246H	Expiry Date	2018-08

Analysis procedure

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	25.0
Master Mix reference	OPTIMA AB Master Mix 4x RNA
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

QC Results by Real Time PCR

Testing	Dye	Specifications	Results
Positive Control (Ct/Cp) : H. Rhinovirus	FAM	30.0 ± 3.3	28.8
Negative Control (Ct/Cp)	N/A	Negative	Negative

Approval

Name (function) : D. KUSINDA (QA OFFICER)	Signature: 	Date: 03/10/2016
---	--	----------------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.