

**Identification of the product**

Product name	R-DiaRhino		
Product reference	DDGR-60-L100	Production date	2017-10
Lot number	2017K27/I	Expiry date	2019-10

**Analysis procedure**

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-R4-A100
Real Time PCR equipment (Supplier)	LightCycler480 II (Roche)

**QC results**

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp) : Rhinovirus	FAM	30.0 ± 3.3	30.2
Negative Control (Ct/Cp)	N/A	Negative	Negative

**Approval**

Name and function	Signature:	Date:
<div style="border: 1px solid black; padding: 2px; display: inline-block;">                 David Kusinda QA Officer             </div>		19/12/2017

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.