

**Identification of the product**

Product name	R-DiaRhino		
Product reference	DDGR-60-L100	Production date	2020-08
Lot number	2020I01/I	Expiry date	2022-08


**Analysis procedure**

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-R4-A100
Real Time PCR equipment (Supplier)	LightCycler 480II (Roche)

**QC results**

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp) :	FAM	30.0 ± 3.3	27.8
Negative Control (Ct/Cp)	N/A	Negative	Negative

**Approval**

Name and function : CARINE COUVREUR QA OFFICER	Signature: 	Date: 11 SEP 2020
------------------------------------------------------	----------------------------------------------------------------------------------------------------	----------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.