

Identification of the product

Product Name	R-DiaTub		
Catalogue Number	DDGR-70-L100	Production Date	2016-11
Lot Number	01006K	Expiry Date	2018-11

Analysis procedure

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-D2-Q100
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

QC Results by Real Time PCR

Testing	Dye	Specifications	Results
Positive Control (Ct/Cp) : <i>Mycobacterium tuberculosis</i>	FAM	30.0 ± 3.3	30.7
Negative Control (Ct/Cp)	N/A	Negative	Negative

Approval

Name (function) : D. KUSINDA QA OFFICER	Signature: 	Date: 08/02/2017
--	---	----------------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.