

Identification of the product

Product name	R-DiaTub		
Product reference	DDGR-70-L100	Production date	2017-09
Lot number	020071	Expiry date	2019-09

Analysis procedure

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-D2-Q100
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

QC results

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp) :	FAM	30.0± 3.3	28.9
Negative Control (Ct/Cp)	N/A	Negative	Negative

Approval

Name and function : <i>R. ZACETELI</i> <i>QA/QC Manager</i>	Signature: 	Date: <i>18/12/2017</i>
---	---	----------------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.