

Identification of the product

Product name	R-DiaTub		
Product reference	DDGR-70-L100	Production date	2018-06
Lot number	03008F	Expiry date	2020-06


Analysis procedure

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-D2-Q100
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

QC results

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp) :	FAM	30.0± 3.3	29.9
Negative Control (Ct/Cp)	N/A	Negative	Negative

Approval

Name and function	Signature:	Date:
<div style="border: 1px solid black; padding: 5px; display: inline-block;"> David Kusinda QA Officer </div>		10/08/2018

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.