

**Identification of the product**

Product name	R-DiaTub		
Product reference	DDGR-70-L100	Production date	2019-01
Lot number	04009A	Expiry date	2021-01

**Analysis procedure**

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-D2-Q100
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

**QC results**

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp) :	FAM	30.0± 3.3	28.3
Negative Control (Ct/Cp)	N/A	Negative	Negative

**Approval**

Name and function : Julianna KERESZTES QA Officer	Signature: 	Date: 20 MAY 2019
---	--	----------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.