

Identification of the product

Product Name	R-DiaHPIV		
Catalogue Number	DDGR-84-L100	Production Date	2016-09
Lot Number	04006I	Expiry Date	2018-09


Analysis procedure

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-R4-A100 (Diagenode)
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

QC Results by Real Time PCR

Testing	Dye	Specifications	Results
Positive Control (Ct/Cp) HPIV 1	FAM	30.0 ± 3.3	FAM : 29.2
Positive Control (Ct/Cp) HPIV 3	YD	30.0 ± 3.3	YD : 30.0
Positive Control (Ct/Cp) HPIV 2	FAM	30.0 ± 3.3	FAM : 30.1
Positive Control (Ct/Cp) HPIV 4	YD	30.0 ± 3.3	YD : 30.2
Negative Control (Ct/Cp)	NA	Negative	Negative

Approval

Name (function) : D. KUSIMEDA QA OFFICER	Signature: 	Date: 15/12/2016
---	--	----------------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.