

Identification of the product

Product name	R-DiaHPIV		
Product reference	DDGR-84-L100	Production date	2017-03
Lot number	05007C	Expiry date	2019-03

Analysis procedure

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-R4-A100
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

QC results

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp) HPiV 1	FAM	30.0 ± 3.3	29.1
Positive Control (Ct/Cp) HPiV 3	YD	30.0 ± 3.3	29.4
Positive Control (Ct/Cp) HPiV 2	FAM	30.0 ± 3.3	29.9
Positive Control (Ct/Cp) HPiV 4	YD	30.0 ± 3.3	30.5
Negative Control (Ct/Cp)	N/A	Negative	Negative

Approval

Name and function : David Kusinda QA Officer	Signature: 	Date: 14/06/2017
--	--	---------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.