

Identification of the product

Product name	R-DiaHPIV		
Product reference	DDGR-84-L100	Production date	2019-06
Lot number	08009F	Expiry date	2021-06


Analysis procedure

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-R4-A100
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

QC results

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp) : HPiV 1	FAM	30 ± 3.3	27.9
Positive Control (Ct/Cp) : HPiV 3	YD	30 ± 3.3	28.8
Positive Control (Ct/Cp) : HPiV 2	FAM	30 ± 3.3	29.1
Positive Control (Ct/Cp) : HPiV 4	YD	30 ± 3.3	29.6
Negative Control (Ct/Cp)	N/A	Negative	Negative

Approval

Name and function : Julianna KERESZTES QA Officer	Signature: 	Date: 09 SEP 2019
---	--	-----------------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.

