

**Identification of the product**

Product name	R-DiaHPIV		
Product reference	DDGR-84-L100	Production date	2019-11
Lot number	09009K	Expiry date	2021-11


**Analysis procedure**

Volume PCR samples (µL)	5.00
Volume PCR mix (µL)	20.00
Master Mix reference	DMML-R4-A100
Real Time PCR equipment (Supplier)	LightCycler480II (Roche)

**QC results**

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp) :HPiV1	FAM	30.0 ± 3.3	28.1
Positive Control (Ct/Cp) :HPiV3	YD	30.0 ± 3.3	28.6
Positive Control (Ct/Cp) :HPiV2	FAM	30.0 ± 3.3	29.3
Positive Control (Ct/Cp) :HPiV4	YD	30.0 ± 3.3	29.5
Negative Control (Ct/Cp)	N/A	Negative	Negative

**Approval**

Name and function : Romy ZALETEL QA Manager	Signature: 	Date: 10/03/2020
---	--	---------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.