

Identification of the product

Product name	R-DiaHPIV		
Product reference	DDGR-84-L100	Production date	2020-08
Lot number	10000H	Expiry date	2022-08


Analysis procedure

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-R4-A100
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

QC results

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp) : HPiV 1	FAM	30.0 ± 3.3	29.3
Positive Control (Ct/Cp) : HPiV 3	YD	30.0 ± 3.3	30.2
Positive Control (Ct/Cp) : HPiV 2	FAM	30.0 ± 3.3	30.6
Positive Control (Ct/Cp) : HPiV 4	YD	30.0 ± 3.3	30.1
Negative Control (Ct/Cp)	N/A	Negative	Negative

Approval

Name and function : <div style="border: 1px solid black; padding: 2px; display: inline-block;"> Dominique d'Amico QA Officer </div>	Signature: 	Date: 28 -10- 2020
---	--	------------------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.