

Identification of the product

Product Name	R-DiaHPIV		
Catalog Number	DDGR-84-L100	Production Date	2016-01
Lot Number	16126A	Expiry Date	2018-01

Analysis procedure

Volume PCR samples (µL)	5
Volume PCR mix (µL)	25
Master Mix reference (Supplier)	OPTIMA AB Master Mix 4x RNA
Real Time PCR equipment (Supplier)	LightCyclr 480 II (Roche)

QC Results by Real Time PCR

Testing	Specifications	Results
Positive Control (Ct/Cp) HPIV 1	30.0 ± 3.3	FAM: 28.3
Positive Control (Ct/Cp) HPIV 3	30.0 ± 3.3	YD:28.4
Positive Control (Ct/Cp) HPIV 2	30.0 ± 3.3	FAM: 28.0
Positive Control (Ct/Cp) HPIV 4	30.0 ± 3.3	YD:29.1
Negative Control (Ct/Cp)	Negative	Negative

Approval

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.

Approved by: R. MALLET  Date: 05 Apr 2016  
*QA/PC Manager*