

Identification of the product

Product name	S-DiaCT		
Product reference	DDGS-10-L100	Production date	2018-10
Lot number	16008J	Expiry date	2020-10


Analysis procedure

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-D2-R100
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

QC results

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp) :	FAM	30.0 ± 3.3	31.1
Negative Control (Ct/Cp)	N/A	Negative	Negative

Approval

Name and function : COUVRUEL E QA OFFICER	Signature: 	Date: 01 FEB 2019
---	--	----------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.