

Identification of the product

Product name	S-DiaCTNG		
Product reference	DDGS-20-L100	Production date	2019-08
Lot number	09009H	Expiry date	2021-08


Analysis procedure

Volume PCR samples (µL)	5
Volume PCR mix (µL)	20
Master Mix reference	DMML-D5-D100
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

QC results

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp)	FAM	27.4 ± 3.3	26.8
	YD	26.1 ± 3.3	25.1
Internal Quality Control 1 (CT)	FAM	10 positive/10 replicates	10 positive
Internal Quality Control 2 (NG)	YD	10 positive/10 replicates	10 positive
Negative Control (Ct/Cp)	N/A	Negative	Negative

Approval

Name and function : S. Lombard QA officer	Signature: 	Date: 08 oct 2019
---	---	----------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.

