

Identification of the product

Product Name	S-DiaCTNG for performance evaluation only		
Catalog Number	DDGS-20-L100	Production Date	2015-01
Lot Number	2015A15/I	Expiry Date	2015-07

Analysis procedure


Volume PCR samples (µL)	5
Volume PCR mix (µL)	25
Master Mix reference (Supplier)	Optima Du Mater Mix x5 DNA (Diagenode)
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

QC Results by Real Time PCR

Testing	Specifications	Results
Positive Control (Ct/Cp) FAM	29.7 ± 3.3	29.8
Positive Control (Ct/Cp) YD	27.5 ± 3.3	28.1
Negative Control (Ct/Cp)	Negative	Negative

Approval

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.

Approved by: R. ZALIEDELS  


Date: 28 Jan 2015