

**Identification of the product**

Product name	S-DiaCTNG		
Product reference	DDGS-20-L100	Production date	2017-11
Lot number	2018C19/I	Expiry date	2019-11



**Analysis procedure**

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-D5-D100
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

**QC results**

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp)	FAM	27.4 ± 3.3	28.5
	YD	26.1 ± 3.3	26.9
Internal Quality Control 1 (CT)	FAM	10 positive/10 replicates	10 positive
Internal Quality Control 2 (NG)	YD	10 positive/10 replicates	10 positive
Negative Control (Ct/Cp)	N/A	Negative	Negative

**Approval**

Name and function:  Dominique d'Amise QA Officer	Signature: 	Date: 17/04/2018
--	---	---------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.