

**Identification of the product**

Product Name	S-DiaGONO		
Catalog Number	DDGS-40-L100	Production Date	2016-03
Lot Number	2016C23/III	Expiry Date	2017-03

**Analysis procedure**

Volume PCR samples (µL)	5
Volume PCR mix (µL)	25
Master Mix reference (Supplier)	OPTIMA Ro Master Mix 2x DNA (Diagenode)
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

**QC Results by Real Time PCR**

Testing	Specifications	Results
Positive Control (Ct/Cp) <i>Neisseria gonorrhoeae</i>	27.0 ± 3.3	FAM : 28.2
Negative Control (Ct/Cp)	Negative	Negative

**Approval**

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.

Approved by: A ZALEVELS  Date: 04 Apr 2016  
*QA / QC Manager*