

Identification of the product

Product Name	S-DiaGono		
Catalog Number	DDGS-40-L100	Production Date	2016-05
Lot Number	2016F16/I	Expiry Date	2017-06

Analysis procedure


Volume PCR samples (µL)	5
Volume PCR mix (µL)	25
Master Mix reference (Supplier)	OPTIMA Ro Master Mix 2x DNA
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

QC Results by Real Time PCR

Testing	Specifications	Results
Positive Control (Ct/Cp) <i>Neisseria gonorrhoeae</i>	27.0 ± 3.3	FAM : 28.5
Negative Control (Ct/Cp)	Negative	Negative

Approval

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.

Approved by: D. KUSINDA  
  
 QA OFFICER

Date: 08/07/2016