

Identification of the product

Product Name	S-DiaGono		
Catalogue Number	DDGS-40-L100	Production Date	2016-09
Lot Number	2016I23/I	Expiry Date	2017-09


Analysis procedure

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	25.0
Master Mix reference	OPTIMA Ro Master Mix 2x DNA
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

QC Results by Real Time PCR

Testing	Dye	Specifications	Results
Positive Control (Ct/Cp) : <i>Neisseria gonorrhoeae</i>	FAM	27.0 ± 3.3	27.9
Negative Control (Ct/Cp)	N/A	Negative	Negative

Approval

Name (function) : D. KUSINDA QA OFFICER	Signature: 	Date: 10/10/2016
--	--	----------------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.