

Identification of the product

Product Name	S-DiaGono		
Catalogue Number	DDGS-40-L100	Production Date	2016-11
Lot Number	2016L05/I	Expiry Date	2017-11


Analysis procedure

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference (Supplier)	DMML-D2-R100 (Diagenode)
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

QC Results by Real Time PCR

Testing	Dye	Specifications	Results
Positive Control (Ct/Cp) : <i>Neisseria meningitidis</i>	FAM	27.0 ± 3.3	27.8
Negative Control (Ct/Cp)	N/A	Negative	Negative

Approval

Name (function) : D. KUSINDA QA OFFICER	Signature: 	Date: 23/12/2016
---	--	---------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.