

**Identification of the product**

Product name	S-DiaGono		
Product reference	DDGS-40-L100	Production date	2020-06
Lot number	2020F08/IV	Expiry date	2022-06


**Analysis procedure**

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-D2-R100
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

**QC results**

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp) :	FAM	27.0 ± 3.3	28.8
Negative Control (Ct/Cp)	N/A	Negative	Negative

**Approval**

Name and function : Caroline TOUSSAINT QA Officer	Signature: 	Date: <b>26 JUN 2020</b>
---	---	-----------------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.