

**Identification of the product**

Product Name	S-DiaMGTV		
Catalogue Number	DDGS-60-L100	Production Date	2016-12
Lot Number	12006L	Expiry Date	2018-12


**Analysis procedure**

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-D5-D100
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

**QC Results by Real Time PCR**

Testing	Dye	Specifications	Results
Positive Control (Ct/Cp) : <i>Mycoplasma genitalium</i>	FAM	30.0 ± 3.3	31.3
Positive Control (Ct/Cp) : <i>Trichomonas vaginalis</i>	YD	30.0 ± 3.3	31.3
Negative Control (Ct/Cp)	N/A	Negative	Negative

**Approval**

<b>Name (function) :</b> D. KUSINDA QA OFFICER	<b>Signature:</b> 	<b>Date:</b> 17/02/2017
--	---	----------------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.