

Identification of the product

Product Name	S-DiaMGTV		
Catalogue Number	DDGS-60-L100	Production Date	2017-03
Lot Number	13007C	Expiry Date	2019-03

Analysis procedure

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-D5-D100
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

QC Results by Real Time PCR

Testing	Dye	Specifications	Results
Positive Control (Ct/Cp) : <i>Mycoplasma genitalium</i>	FAM	30.0 ± 3.3	31.2
Positive Control (Ct/Cp) : <i>Trichomonas vaginalis</i>	YD	30.0 ± 3.3	30.9
Negative Control (Ct/Cp)	N/A	Negative	Negative

Approval

Name (function) :	David Kusinda QA Officer	Signature:		Date:	27/06/2017
--------------------------	-----------------------------	-------------------	--	--------------	------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.