

Identification of the product

Product name	S-DiaMGTV		
Product reference	DDGS-60-L100	Production date	2017-10
Lot number	14007J	Expiry date	2019-10


Analysis procedure

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-D5-D100
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

QC results

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp) : <i>Mycoplasma genitalium</i>	FAM	30.0 ± 3.3	31.1
Positive Control (Ct/Cp) : <i>Trichomonas vaginalis</i>	YD	30.0 ± 3.3	29.9
Negative Control (Ct/Cp)	N/A	Negative	Negative

Approval

Name and function	Signature:	Date:
David Kusinda QA Officer		10/01/2018

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.