

**Identification of the product**

Product name	S-DiaMGTV		
Product reference	DDGS-60-L100	Production date	2018-08
Lot number	16008H	Expiry date	2020-08


**Analysis procedure**

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-D5-D100
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

**QC results**

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp) : <i>Mycoplasma genitalium</i>	FAM	30.0 ± 3.3	30.0
Positive Control (Ct/Cp) : <i>Trichomonas vaginalis</i>	YD	30.0 ± 3.3	29.2
Negative Control (Ct/Cp)	N/A	Negative	Negative

**Approval**

Name and function :	Signature:	Date:
Dominique d'Amico QA Officer		22/11/2018

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.