

Identification of the product

Product name	S-DiaMGTV		
Product reference	DDGS-60-L100	Production date	2020-06
Lot number	19000F	Expiry date	2022-06

Analysis procedure

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-D5-D100
Real Time PCR equipment (Supplier)	LightCycler 480II (Roche)

QC results

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp) :	FAM	30.0 ± 3.3	30.8
Positive Control (Ct/Cp) :	YD	30.0 ± 3.3	29.9
Negative Control (Ct/Cp)	N/A	Negative	Negative

Approval

Name and function : Carine COUVRUER QA Officer.	Signature: 	Date: 18 SEP 2020.
---	--	-----------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.