

**Identification of the product**

Product name	S-DiaMGRes		
Product reference	DDGS-65-L025	Production date	2019-02
Lot number	2019B12/II	Expiry date	2020-02



**Analysis procedure**

Volume PCR samples (µL)	Plasmid and UIC: 5.0+2.5
Volume PCR mix (µL)	17.5
Master Mix reference	N/A
Real Time PCR equipment (Supplier)	ABI 7500 FAST (Applied Biosystems)

**QC results**

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp) :	FAM	30.0 ± 3.3	30.0
Positive Control (Ct/Cp) :	YD	30.0 ± 3.3	29.1
Positive Control (Ct/Cp) :	ROX	30.0 ± 3.3	30.2
Internal Control (Ct/Cp) : UIC	CY5	30.0 ± 3.3	30.2
Negative Control (Ct/Cp)	N/A	Negative	Negative

**Approval**

Name and function :	Signature:	Date:
 Dominique d'Amico QA Officer		20 - 03 - 20 - 13

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.