

Identification of the product

Product name	S-DiaMGRes		
Product reference	DDGS-65-L025	Production date	2020-03
Lot number	2020E08/1	Expiry date	2021-03

Analysis procedure

Volume PCR samples (µL)	Plasmid and UIC: 5.0+2.5
Volume PCR mix (µL)	17.5
Master Mix reference	N/A
Real Time PCR equipment (Supplier)	ABI 7500 FAST (Applied Biosystems)

QC results

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp) :	FAM	30.0 ± 3.3	28.9
Positive Control (Ct/Cp) :	YD	30.0 ± 3.3	28.3
Positive Control (Ct/Cp) :	ROX	30.0 ± 3.3	30.0
Internal Control (Ct/Cp) : UIC	CY5	30.0 ± 3.3	31.9
Negative Control (Ct/Cp)	N/A	Negative	Negative

Approval

Name and function : Carine COUVRUER QA Officer	Signature: 	Date: 05 JUN 2020
--	--	----------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.