

Identification of the product

Product name	S-DiaMGRes		
Product reference	DDGS-65-L025	Production date	2020-08
Lot number	2020H28/1	Expiry date	2021-08

Analysis procedure

Volume PCR samples (µL)	Plasmid and UIC: 5.0 +2.5
Volume PCR mix (µL)	17.5
Master Mix reference	N/A
Real Time PCR equipment (Supplier)	ABI 7500 FAST (Applied Biosystems)

QC results

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp) :	FAM	30.0 ± 3.3	29.1
Positive Control (Ct/Cp) :	YD	30.0± 3.3	30.2
Positive Control (Ct/Cp) :	ROX	30.0± 3.3	30.3
Positive Control (Ct/Cp) : UIC	CY5	30.0± 3.3	29.4
Negative Control (Ct/Cp)	N/A	Negative	Negative

Approval

Name and function : Caroline TOUSSAINT QA Officer	Signature: 	Date: 28 SEP 2020
---	--	-----------------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.