

Certificate of Analysis

Product identification			
Product name	S-DiaMGRes		
Product reference	DDGS-65-L025	Manufacturing date	2021-06
Lot number	2021J04/I	Expiry date	2023-06

Analysis procedure	
Volume PCR samples (µL)	Plasmid and UIC: 5.0+2.5
Volume PCR mix (µL)	17.5
Master Mix reference	N/A
Real Time PCR equipment (Supplier)	ABI 7500 FAST (Applied Biosystems)

QC Results				
Testing	Channel	Specification	Results	Status
Positive Control (Ct/Cp):	FAM	26.7-33.3	30.2	Passed / Fail
Positive Control (Ct/Cp):	YD	26.7-33.3	29.6	Passed / Fail
Positive Control (Ct/Cp):	ROX	26.7-33.3	30.4	Passed / Fail
Positive Control (Ct/Cp): UIC	CY5	26.7-33.3	30.2	Passed / Fail
Negative Control (Ct/Cp)	N/A	Negative	Negative	Passed / Fail

Approval			
	Name & Function	Signature	Date
QC	Philippe Naud Je Spécialiste		20-10-2021
QA	LAMARI Samira QA officer		27 OCT 2021

This batch has been released by Quality Assurance in compliance with Diagenode's Quality Management System.

