


Product identification			
Product name	R-DiaAdeno		
Product reference	DPSR-15-L080	Production date	2020-05
Lot number	04000E	Expiry date	2021-08

Analysis procedure	
Real Time PCR equipment (Supplier)	GeneLEADVIII (Precision System Science Co Ltd (PSS))

QC results			
Testing	Channel	Specifications	Results
Positive Control (Ct/Cp)	FAM	30.1 ± 3.3	30.0
IQC (Ct/Cp)	FAM	31.4 ± 3.3	31.7
Negative Control (Ct/Cp)	N/A	Negative	Negative
EIC Control (Ct/Cp)	Cy5	28.7 ± 3.3	30.0

Approval		
Name :	Signature:	Date:
<b>Julianna KERESZTES</b>		<b>03 DEC 2020</b>
Function : <b>QA Officer</b>		

This lot has been reviewed by Quality Assurance in compliance of Diagenode Diagnostics Quality Management System.