

Product identification			
Product name	R-DiaPnJ		
Product reference	DPSR-20-L040	Production date	2020-01
Lot number	2020F02/II	Expiry date	2021-05

Analysis procedure			
QC results on Roche LC480			
Testing	Channel	Specifications	Results
Positive Control (Ct/Cp)	FAM	$28.9 \leq Ct \leq 36.0$	33.7
Positive Control (Ct/Cp)	HEX	$24.9 \leq Ct \leq 32.4$	28.1
IQC (Ct/Cp)	FAM	$30.4 \leq Ct \leq 36.8$	35.5
IQC (Ct/Cp)	HEX	$28.8 \leq Ct \leq 34.3$	32.8
Negative Control (Ct/Cp)	N/A	Negative	Negative
EIC Control (Ct/Cp)	Cy5	$\leq 36.0$	30.8
QC results on Genelead VIII			
Testing	Channel	Specifications	Results
Positive Control (Ct/Cp)	FAM	$30.9 \leq Ct \leq 33.7$	32.7
Positive Control (Ct/Cp)	HEX	$27.6 \leq Ct \leq 30.2$	29.5
IQC (Ct/Cp)	FAM	$31.4 \leq Ct \leq 35.4$	34.4
IQC (Ct/Cp)	HEX	$30.5 \leq Ct \leq 33.8$	33.2
Negative Control (Ct/Cp)	N/A	Negative	Negative
EIC Control (Ct/Cp)	Cy5	$\leq 36.0$	30.0

Quality Control: Laszlo Mahacsik Date: 24.07.2020

Quality Assurance: Julianna KERESZTES Date: 24 JUL 2020  
QA Officer

Comments:

N/A