


Product identification			
Product name	R-DiaFluRSV		
Product reference	DPSR-35-L080	Production date	2020-01-15
Lot number	2019K25/II	Expiry date	2021-04

Analysis procedure	
Volume PCR samples (µL)	10.0
Volume PCR mix (µL)	30.0
Real Time PCR equipment (Supplier)	geneLEADVIII (Precision System Science Co Ltd (PSS))

QC results			
Testing	Channel	Specifications	Results
Positive Control (Ct/Cp)	FAM	30.0 ± 3.3	31.2
	HEX	30.0 ± 3.3	30.9
	ROX	31.0 ± 3.3	31.4
Negative Control (Ct/Cp)	N/A	Negative	Negative
Internal Control (Ct/Cp)	Cy5	29.0 ± 3.3	28.9

Approval		
Name :	Signature:	Date:
Caroline TOUSSAINT		12 MAR 2020
Function : QA Officer		

This lot has been reviewed by Quality Assurance in compliance of Diagenode Diagnostics Quality Management System.