


Product identification			
Product name	R-DiaMTBComplex		
Product reference	DPSR-70-L100	Production date	2020-03
Lot number	2020F02/III	Expiry date	2021-03

Analysis procedure	
Volume PCR samples (µL)	10.0
Volume PCR mix (µL)	30.0
Real Time PCR equipment (Supplier)	geneLEADVIII (Precision System Science Co Ltd (PSS))

QC results			
Testing	Channel	Specifications	Results
Positive Control (Ct/Cp)	FAM	30.0 ± 3.3	29.0
IQC (Ct/Cp)	FAM	36.0 ± 3.3	34.7
Negative Control (Ct/Cp)	N/A	Negative	Negative
Internal Control (Ct/Cp)	Cy5	27.0 ± 3.3	27.1

Approval		
Name :	Signature:	Date:
Caroline TOUSSAINT		16 JUL 2020
Function : QA Officer		

This lot has been reviewed by Quality Assurance in compliance of Diagenode Diagnostics Quality Management System.