

Identification of the product

Product name	G-DiaDiff		
Product reference	DDGG-30-L100	Production date	2019-04
Lot number	2019D16/I	Expiry date	2020-04

Analysis procedure

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-D2-Q100
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

QC results

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp) : <i>Clostridium difficile tcdA</i>	FAM	31.0 ± 3.3	29.7
Positive Control (Ct/Cp) : <i>Clostridium difficile tcdB</i>	YD	30.0 ± 3.3	28.5
Negative Control (Ct/Cp)	N/A	Negative	Negative

Approval

Name and function : Carine COUVRUER QA Officer	Signature: 	Date: 03 MAY 19
--	--	------------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.