

Identification of the product

Product name	G-DiaDiff		
Product reference	DDGG-30-L100	Production date	2020-08
Lot number	2020H14/I	Expiry date	2022-08

Analysis procedure

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-D2-Q100
Real Time PCR equipment (Supplier)	LightCycler480II (Roche)

QC results

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp) :	FAM	31.0 ± 3.3	31.5
Positive Control (Ct/Cp) :	YD	30.0 ± 3.3	29.7
Negative Control (Ct/Cp)	N/A	Negative	Negative

Approval

Name and function : Caroline TOUSSAINT QA Officer	Signature: 	Date: 27 AUG 2020
---	--	-----------------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.