

**Identification of the product**

Product Name	G-DiaNota		
Catalog Number	DDGG-10-L100	Production Date	2015-12
Lot Number	04005L	Expiry Date	2017-12

**Analysis procedure**


Volume PCR samples (µL)	5
Volume PCR mix (µL)	25
Master Mix reference (Supplier)	Optima PLUS Master Mix (Diagenode)
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

**QC Results by Real Time PCR**

Testing	Specifications	Results
Positive Control (Ct/Cp) <i>Norovirus I</i>	28.0 ± 3.3	FAM : 27.1
Positive Control (Ct/Cp) <i>Norovirus II</i>	30.0 ± 3.3	YD : 30.8
Positive Control (Ct/Cp) <i>Rotavirus</i>	28.0 ± 3.3	TR : 27.3
Internal Control (Ct/Cp) EIC	≤ 36.0	Cy5 : 30.9
Negative Control (Ct/Cp)	Negative	Negative

**Approval**

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.

Approved by:   
*QA/PC Manager*

Date: 07 Nov 2016