

Product identification			
Product name	H-DiaCMVQ		
Product reference	DDGH-10-L200	Production date	2019-05
Batch number	03009E	Expiry date	2020-05

Analysis procedure	
Volume PCR samples (µL)	Positive control: Standards:
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-D2-Q100
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

QC results			
Testing	Channel	Specifications	Results
Positive Control (Ct/Cp)	FAM	21.2—24.2	23.1
IQC (c/PCR)	FAM	222—1802	502
EIC DNA (Ct)	CY5	< 36.0	26.0
Standard 1 (Ct/Cp)	FAM	17.9—20.9	19.7
Standard 2 (Ct/Cp)	FAM	21.2—24.2	22.9
Standard 3 (Ct/Cp)	FAM	24.5—27.5	26.3
Standard 4 (Ct/Cp)	FAM	27.8—30.8	29.8
Standard 5 (Ct/Cp)	FAM	31.1—34.1	33.1
Negative Control (Ct/Cp)	N/A	Negative	Negative
Slope	N/A	-3.6 to -3.0	-3.4
Efficiency	N/A	1.9 to 2.1	2.0

Quality Control: CHARLIER *Charlier* Date: 06/08/2019

Quality Assurance: Dominique d'Amico  
QA Officer Date: 06-08-2019