


Product identification			
Product name	H-DiaEBVQ		
Product reference	DDGH-20-L200	Production date	2019-11
Lot number	2019I24/V	Expiry date	2020-11

Analysis procedure	
Volume PCR samples (µL)	5
Volume PCR mix (µL)	20
Master Mix reference	DMML-D2-Q100
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

QC results				
Testing		Channel	Specifications	Results
Positive Control (Ct/Cp)		FAM	21.5 - 24.5	23.2
Production Quality Control (Plasma)	(c/PCR)	FAM	200 - 2000	780
Standard 1 (Ct/Cp)		FAM	18.2 - 21.2	19.7
Standard 2 (Ct/Cp)		FAM	21.5 - 24.5	23.0
Standard 3 (Ct/Cp)		FAM	24.8 - 27.8	26.2
Standard 4 (Ct/Cp)		FAM	28.1 - 31.1	29.8
Standard 5 (Ct/Cp)		FAM	31.4 - 34.4	32.9
Negative Control (Ct/Cp)		N/A	Negative	Negative
Slope		N/A	-3.6 to -3.0	-3.3
Efficiency		N/A	1.9 to 2.1	2.01

Quality Control: CHARLIER France  Date: 02/03/2020

Quality Assurance: Dominique d'Amico
QA Officer  Date: 11-03-2020

This lot has been reviewed by Quality Assurance in compliance of Diagenode Diagnostics Quality Management System