

Identification of the product

| | | | |
|-------------------|--------------|-----------------|---------|
| Product name | O-DiaADV | | |
| Product reference | DDGO-10-L100 | Production date | 2018-05 |
| Lot number | 03008E | Expiry date | 2020-05 |


Analysis procedure

| | |
|------------------------------------|----------------------------|
| Volume PCR samples (µL) | 5.0 |
| Volume PCR mix (µL) | 20.0 |
| Master Mix reference | DMML-D2-D600 |
| Real Time PCR equipment (Supplier) | LightCycler 480 II (Roche) |

QC results

| Testing | Channel | Specifications | Results |
|----------------------------|---------|----------------|----------|
| Positive Control (Ct/Cp) : | FAM | 30.0 ± 3.3 | 28.2 |
| Negative Control (Ct/Cp) | N/A | Negative | Negative |

Approval

| | | |
|---|--|---------------------|
| Name and function : <div style="border: 1px solid black; padding: 2px; display: inline-block;"> Dominique d'Amico QA Officer </div> | Signature:  | Date: 06/08/2018 |
|---|--|---------------------|

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.