

Identification of the product

Product Name	R-DiaFlu		
Catalog Number	DDGR-30-L100	Production Date	2014-08
Lot Number	02004H	Expiry Date	2016-08

Analysis procedure


Volume PCR samples (µL)	5
Volume PCR mix (µL)	25
Master Mix reference (Supplier)	TaqMan Fast Virus 1-Step Master Mix (Applied Biosystems)
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

QC Results by Real Time PCR

Testing	Specifications	Results
Positive Control (Ct/Cp)	FAM : 30.0 ± 3.3	FAM : 28.6
Positive Control (Ct/Cp)	YD : 30.0 ± 3.3	YD : 30.4
Negative Control (Ct/Cp)	Negative	Negative

Approval

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.

Approved by: F. Noye QA office


Date: 18 Jan 2016